Organisation name Site or unit		FORM Anaesthesia machine and equipment check	
Document ID	SRG-FRM-003	**Version**	1.0
Status	Blank form, to adapt	**Issue date**	
Review date		**Replaces**	
Owner		**Approved by**	
Signature			

Form · SRG-FRM-003 · Version 1.0 · Approved by the ASF Expert Committee

Supports: 2.2

1. Purpose

The daily check of the anaesthetic machine, monitors, suction and emergency airway equipment before the first case.

2. How to complete

- Complete before the first case of each list.
- Report any fault and take the machine out of use.
- Sign the check.

3. Checks

Check	Entry
Machine leak test passed	Yes
Oxygen supply and reserve cylinder checked	Yes
Monitors switched on and alarms tested	Yes
Suction working	Yes
Airway equipment and difficult airway trolley checked	Yes

Drugs and emergency drugs checked	Yes
Checked by and time	Yes

4. Evidence

Reference	Note
E-009	Based on the ASA Standards for Basic Anesthetic Monitoring (2020).
E-006	ASF Policy Framework v1.2. Adapted.
E-009	ASF Expert Committee consensus.

5. Change history

Version	Date	What changed	Approval
1.0	the date of approval	First version	set Expert Committee approval

6. References

- Accréditation Sans Frontières (2026). ASF Policy Framework v1.2. (E-006)

- Accréditation Sans Frontières (2026). ASF Expert Committee consensus on Standard 1. (E-009)

Download this document as PDF